



CERTIFICATE OF COMPLIANCE

LLOYD'S REGISTER QUALITY ASSURANCE

hereby declares that the certification assessment has demonstrated that

**RiboFlow® *Listeria Twin* Detection Kit (orderno. SY-LAB 51-419113),
after two enrichment steps in ONE-broth *Listeria* (BiMedia 404A)
(orderno. SY-LAB 41-472200)**

Manufactured and supplied by:
SY-LAB Geräte GmbH.
Tullnerbachstraße 61-65
3011 Neupurkersdorf
Austria

has been validated for the detection of *Listeria monocytogenes* and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website: www.microval.org

Reference method: EN ISO 11290-1:2017; Microbiology of the food chain — Horizontal method for the detection and enumeration of *Listeria monocytogenes* and of *Listeria spp.* — Part 1: Detection method.

Scope: raw milk and dairy products, heat-processed milk and dairy products, ready-to-eat, ready-to-reheat meat products and ready-to-eat, ready-to-reheat fishery products.

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules and Certification Scheme version 8.

Certificate no.: 2015LR53

First approval date: 22 November 2018
Expiry date: 21 November 2022



ISSUED BY: Lloyd's Register Nederland B.V.
Rotterdam, The Netherlands